



Republic of the Philippines
DEPARTMENT OF AGRICULTURE
 Regional Field Office 13
 Capitol Site Butuan City 8600

September 23, 2024

SUPPLEMENTAL BID BULLETIN NO. 2024- 057

This is to inform **ALL PROSPECTIVE BIDDERS** that the Invitation to Bid posted last September 12, 2024 for the Procurement of Various Laboratory Equipment under Reference No. 11240281 with a total ABC of **Two Million Eight Hundred Thousand Pesos Only (₱ 2,800,000.00)** has the following amendments:

Bid Document	Supplemental
<ul style="list-style-type: none"> I. Bid Data Sheet ITB Clause 20.2 Latest Tax Payments (Monthly/Quarterly) Certificate of Stock Availability Brochure/Product Catalogue II. Section VI. Schedule of Requirement Delivery is 90 days upon issuance of Notice to Proceed III. Section VII. Technical Specifications: <i>Item No. 1 Analytical Balance</i> <u>Specifications:</u> -Response Time: ≤ 2.5 sec. -Calibration Certificate : Yes <i>Item No. 2 Fumehood</i> <u>Specifications:</u> <ul style="list-style-type: none"> - Dimension (WxDxH): $50\pm 5'' \times 25\pm 5'' \times 50\pm 5''$ - Standards Conformity: CFR 29, Part 1910; SEFA 1-2020; NFPA 45-2011; ASHRAE 110-1995; ANSI Z9.5-2011; UL 61010-1; CAN/CSA C22.2 No. 61010.1; UL 1805; SEFA 8-2010, Cabinet Surface Finish Test; CE Conformity Marking <i>Item No. 3 Fumehood</i> <u>Specifications:</u> -Acid Storage Cabinet: (Bench-top Only) 	<ul style="list-style-type: none"> I. Bid Data Sheet ITB Clause 20.2 Latest Tax Payments (Monthly/Quarterly) Certificate of Stock Availability Brochure/Product Catalogue Certificate of Authorized Distributorship II. Section VI. Schedule of Requirement Delivery is 120 days upon issuance of Notice to Proceed III. Section VII. Technical Specifications: <i>Item No. 1 Analytical Balance</i> <u>Specifications:</u> -Response Time: ≤ 3.0 sec. Calibration Certificate : Yes (By 3rd Party) <i>Item No. 2 Fumehood</i> <u>Specifications:</u> <ul style="list-style-type: none"> -Dimension (WxDxH): Maximum of $47'' \times 32'' \times 55''$ - Standards Conformity: CFR 29, Part 1910; SEFA 1-2020; NFPA 45-2011; ASHRAE 110-1995; ANSI Z9.5-2011; UL 61010-1; CAN/CSA C22.2 No. 61010.1; UL 1805; SEFA 8-2010, Cabinet Surface Finish Test; CE Conformity Marking; or their corresponding equivalents <i>Item No. 3 Fumehood</i> <u>Specifications:</u> - Acid Storage Cabinet: None (Bench-top Only)

<p>- Dimension (WxDxH): 50±5" x 25±5" x 50±5"</p> <p>-Standards Conformity: CFR 29, Part 1910; SEFA 1-2020; NFPA 45-2011' ASHRAE 110-1995; ANSI Z9.5-2011; UL 61010-1; CAN/CSA C22.2 No. 61010.1; UL 1805; SEFA 8-2010, Cabinet Surface Finish Test; CE Conformity Marking</p> <p>- Inclusion: 1 pc laboratory stool</p>	<p>-Dimension (WxDxH): Maximum of 47" x 32" x 55"</p> <p>- Standards Conformity: CFR 29, Part 1910; SEFA 1-2020; NFPA 45-2011; ASHRAE 110-1995; ANSI Z9.5-2011; UL 61010-1; CAN/CSA C22.2 No. 61010.1; UL 1805; SEFA 8-2010, Cabinet Surface Finish Test; CE Conformity Marking; or their corresponding equivalents</p> <p>-Inclusion: 1 pc laboratory stool (stainless steel, heavy duty, with foot ring, adjustable height, 360° rotation)</p>
--	--

This supplemental bid bulletin shall be posted at the DA-Caraga and PhilGEPS official websites and at the Office Bulletin Board for the information and guidance of all interested parties and shall form as an integral part of the Bid Documents for the said project.

For: 
ABELLE WAGAS
 BAC Chairperson 